

Minsk republic electric power enterprise Minskenergo (RUE "Minskenergo")

Slutsk Electric Networks 223610, Minsk region, Slutsk, Energetikov St., 1 Tel.: (+375 1795) 7-19-59, fax: (+375 1795) 6-11-23

Director: Anatoly V. Khramtsevich Chief Engineer: Sergey G. Shmelkov

Review to MIKO-8M(A) milliohmmeter

We have been operating MIKO-8M(A) milliohmmeter for two years, and the advantages of the device are the following:

- portability;
- compact size;
- informative menu;
- availability of the DRM test;
- possibility of use on high-inductive equipment.

The disadvantage is a small internal memory for information storing.

MIKO-8M(A) milliohmmeter deserves a high rating according to its characteristics. We are looking forward to new developments.

Для Head of ISPS

1

A.S. Baturin