

SVERDLOVSKY PLANT OF CURRENT TRANSFORMERS

Cherkasskaya Str., 25, Ekaterinburg, 620043; Tel. (343) 212-52-55; 232-64-00 www.cztt.ru; e-mail: cztt@cztt.ru

Dtd. 17/05/2012 # 07-89

ATTN: T.O.Ekaterinina, Head of Marketing and Sales Department,
Lermontov Str., 130, Irkutsk, 664033, P.O. Box 407; Tel./Fax (3952) 42-89-21, 42-34-38

Dear Madam,

To your request #152 dated 14.05.2012 please be informed that:

- 1) Our Company used MIKO-2.3 instruments #18 manufactured in 2009, and #18 manufactured in 2010.
- 2) The instruments are used for measuring the resistance of main switchgears, circuits, of current and power transformers.
- 3) The instruments are rather convenient in operation though length of conductors is sometimes insufficient for measuring the switchgear resistance.
- 4) We failed to use the instrument for measuring the resistance of transformer windings during thermal tests (in the conditions of rapid change in the windings temperature).
- 5) The MIKO-2.3 can be recommended for use by other companies.

Head of TsZI

/Signature/

Ershov S.V.

Prepared by: A.V.Vedemikov,
Tel./Fax: (343) 232 59 95
E-mail: czl@cztt.ru

CERTIFICATE OF TRANSLATION

This is to certify that the foregoing translation from Russian into English was made by the undersigned translator Galliamova Vera of the City of Irkutsk, Russia who is conversant with both these languages, and that, to the best of my knowledge and understanding, it is a true and complete rendition of the corresponding original document.

"Association of translation and interpreters" Plc.

[Seal: The Russian Federation; Irkutsk; Public limited company; "Association of translation and interpreters"; Primary state registration number 1073808003961; TIN 3808162520]

Address: office 208, Sukhe-Batora str., 7, Irkutsk, 664011, Russia; tel.: +7 (3952) 588-722

Signed: _____

Sworn before me at the City of Irkutsk, this 29th day of July, 2015.





№ 07-89 от «17» 05 2012г.

Начальнику отдела сбыта и маркетинга
СКБ ЭП
Т. О. Екатериной

664033, г. Иркутск, ул. Лермонтова, 130, а/я 407
Тел./факс (3952) 42 89 21, 42 34 38

Уважаемая г-жа Екатерина!

На Ваш запрос № 152 от 14.05.2012 г. сообщаем:

- 1) На нашем предприятии применяются приборы МИКО-2.3 №18 2019г. и №18 2010 г.
- 2) Приборами измеряется сопротивление главных цепей КСО, КРУ, трансформаторов тока и силовых трансформаторов.
- 3) Приборы достаточно удобны в использовании. Иногда не хватает длины проводников при измерении сопротивления КРУ.
- 4) Не удалось использовать прибор для измерения сопротивления обмоток трансформаторов при испытаниях на нагрев (в условиях быстрого изменения температуры обмоток).
- 5) Мы рекомендуем МИКО-2.3 для использования другими организациями.

Начальник ЦЗЛ



Ершов С. В.

Исполнитель:
Ведерников А.В.
Тел./Факс (343) 232 59 95
E-mail czl@cztz.ru